



ಕರ್ನಾಟಕ ಸರ್ಕಾರ

ಬೆಳಗಾವಿ ವೈದ್ಯಕೀಯ ವಿಜ್ಞಾನ ಸಂಸ್ಥೆ, ಬೆಳಗಾವಿ

(ಸ್ವಾಯತ್ತ ಸಂಸ್ಥೆ)

GOVERNMENT OF KARNATAKA

BELAGAVI INSTITUTE OF MEDICAL SCIENCES

Dr. B.R. Ambedkar Road, Belagavi -590 001.

(Autonomous Institution)

(Hospital) Phone No: 0831 2420320, (College) Phone No.: 0831 2421464, (College) Fax No.: 0831 2403126

No.BIMS/Hosp/Sup-/CR/76 /2021-22

ದಿನಾಂಕ:

ಕೊಟೇಶನ್ ಅಧಿಸೂಚನೆ

ವಿಷಯ: ಬಿಮ್ಸ್ ಆಸ್ಪತ್ರೆಯಲ್ಲಿ ಕೋವಿಡ್-19 ಚಿಕಿತ್ಸೆಗಾಗಿ ಉಪಯೋಗಿಸಲು ಬೇಕಾಗಿರುವ ಉಪಕರಣಗಳನ್ನು ಸರಬರಾಜು ಪಡೆಯುವ ಬಗ್ಗೆ.

ಬಿಮ್ಸ್ ಆಸ್ಪತ್ರೆಯಲ್ಲಿ ಕೋವಿಡ್-19 ಚಿಕಿತ್ಸೆಗಾಗಿ ಉಪಯೋಗಿಸಲು ಅತೀ ಅವಶ್ಯಕವಾಗಿ ಬೇಕಾಗಿರುವ ಉಪಕರಣಗಳನ್ನು ಖರೀದಿಸಲು ಪ್ರತಿಷ್ಠಿತ ತಯಾರಕರು ಅಥವಾ ಅಧಿಕೃತ ಮಾರಾಟಗಾರರಿಂದ ದರಗಳನ್ನು ಆಹ್ವಾನಿಸಲಾಗಿದೆ. ಆಸಕ್ತಿಯುಳ್ಳವರು ಈ ಕೆಳಗಿನ ಷರತ್ತಿಗೊಳಪಟ್ಟು ತಮ್ಮ ದರಗಳನ್ನು ಈ ಅಧಿಸೂಚನೆ ಜೊತೆಗೆ ಲಗತ್ತಿಸಿದ ನಮೂನೆ (Format of Quotation) ಯಲ್ಲಿ ದಿನಾಂಕ: ----- ರಂದು ಸಾಯಂಕಾಲ 4.00 ಗಂಟೆಯ ಒಳಗಾಗಿ ಕಛೇರಿಗೆ ತಲುಪುವಂತೆ ಸಲ್ಲಿಸತಕ್ಕದ್ದು.

Sl. No.	Name of the Item	Price per unit to be Quoted
1	Fumigation Machine <u>Technical Specification :</u> <ul style="list-style-type: none"> Uniform spread and sufficient contact time with micro organisms to achieve max. Disinfection. Precision metering valve system. Least possible wetness on floor. Easy to clean, detachable, non-corrosive. High grade chemical solution tank with 5 litres capacity. Specially designed intake air filter to prevent dust/dirt and chemicals which helps to protect the motor. Sterile areas ready within 1-1 ½ hr. max. Consistent particle size generation : Submicron (non-wetting) Reach : 20-30 Ft. & 15-20 ft. height cover Space treatment: Up to 7,000-10,000 cu. ft. Chemical fogged: All water based disinfectants, compatible with all leading disinfectant / fumigant solutions. Motor: 20,000 RPM, double stage with tapered fans. Input power: 220V AC, 5-6 Amp., 50 Hz. (Power cable length – 2.5m, 5m optional) Nozzle Assembly: Non-rotating, Vortex Design, Non Clogging, MOC: Engg. Plastic Power Head: High grade Engg. Plastic, unbreakable & shock proof. Solution Tank: 5 litres (MOC-HDPE) with gradual marking / SS 304 Tank available (optional). Strainer / Handle / Mounting Ring: Stainless Steel 304. Precision Metering Systems: Discharge capacity 0-100 ml. / min. Intake Air Filter: Suitable design with double layers for primary & secondary filtration to prevent entry of dust/ dirt & moisture from the air. Empty Weight: 5.5kg. (with plastic tank) / 7kg. (with SS tank) 	3
2	Suction Apparatus <u>Technical Specification :</u> <ul style="list-style-type: none"> Should meet all surgical fields of applications. Should run on oil immersed rotary vane vacuum pump and thermally protected vacuum pump motor. Should have stainless steel cabinet mounted on antistatic caster wheels. Unit should be provided with autoclavable / unbreakable collection jars made from high quality polycarbonate material. The collection jars should have self sealing rubber lids. Vacuum should be adjusted with the help of vacuum regulator and vacuum gauge. Should prevent entry of suction fluid in the pump in case of overfilling. Should have a special bacteria filter to ensure zero contamination and thus preventing infection. Vacuum: 710 mm of Hg at sea level Capacity: 40-50 LPM Power: 0.25-0.5 HP Noise level: ≤50dB ± 3 Power Supply: 220/230V AC Collection Jars: 1.5 liters (2 nos) Mains Cable: 2.5 meters Exhaust fan (optional) Dimensions (mm): 490x315x730 	3
3	ECG Machine 3 Channel <u>Technical Specification :</u> <ul style="list-style-type: none"> Digital 3 Channel ECG machine. 3 Channel real time print width of paper 80mm Simultaneous 12 lead ECG Signal acquisition 4.3 inch color graphic LCD Display Light weight, easy to carry, size of Au Paper Easy to operate with quick color keys Build in rechargeable battery Analysis with interpretation 	3

4	<p>B P Apparatus Digital</p> <p>Technical Specification:</p> <ul style="list-style-type: none">• The digital BP monitor should be CE certified and the manufacturers should be ISO certified company.• Display: LCD digital display• Measurement: Osillometric method• Measurement Range: Pressure 0-270 mmHg and Pulse: 50-170 beats/min• Accuracy: Pressure +3 or -3mm Hg. Pulse rate: +5 or -5% of reading• Should be able to store 100 or more memories in groups• The monitor should indicate irregular heart beat.• Operating temperature/Humidity: +5° C to 40°C• Storage temperature: -25°C to +70°C• Power savings: Automatic power off if system is unused for 3 minutes.• Electric safety Requirements: Should be tested as per relevant standards• Should work on battery and optional provision should be available for operating on mains through adapter.• The weight of the device should not be more than 300gm excluding battery.• The main unit to be supplied with Adult Arm Cuff, battery and instruction manual.	10
---	---	----

ಸೂಚನೆ: ಅಗತ್ಯಬಿದ್ದರೆ ಮೇಲ್ಕಾಣಿಸಿದ ಉಪಕರಣಗಳ ಪದಾರ್ಥನಕ್ಕಾಗಿ ಕರೆಯಲಾಗುವುದು.

1. ತಾವು ನಮೂದಿಸಿದ ದರವು ಯಾವುದೇ ಕಾರಣಕ್ಕೂ ಮಾರುಕಟ್ಟೆ ದರಕ್ಕಿಂತಲೂ ಹೆಚ್ಚಿಗೆಯಾಗಿರಕೂಡದು
2. ಸದರಿ ದರಗಳನ್ನು ಯಾವುದೇ ಕಾರಣ ನೀಡದೆ ತಿರಸ್ಕರಿಸುವ ಹಕ್ಕನ್ನು ನಿರ್ದೇಶಕರು, ಬಿಮ್ಸ್, ಬೆಳಗಾವಿ ಇವರಿಗೆ ಕಾಯ್ದಿರಿಸಲಾಗಿದೆ.
3. ಸದರಿ ವಸ್ತುಗಳ ಖರೀದಿಯು ಸಾರ್ವಜನಿಕ ಸಂಗ್ರಹಣೆಗಳಲ್ಲಿ ಪಾರದರ್ಶಕತೆ ನಿಯಮ 1999ರ ಮತ್ತು 2000 ನಿಯಮಗಳಿಗೆ ಒಳಪಟ್ಟಿರುತ್ತದೆ.
4. ದರಗಳ ಜೊತೆಗೆ GST ನೊಂದಾಯಿತ ದೃಢೀಕರಣ ಪತ್ರ, ಹಾಗೂ ಪ್ಯಾನ್ ಕಾರ್ಡ್ (PAN Card) ನಕಲು ಪ್ರತಿಗಳನ್ನು ಕಡ್ಡಾಯವಾಗಿ ಲಗತ್ತಿಸತಕ್ಕದ್ದು.
5. ಮಾರುಕಟ್ಟೆ ದರ (MRP) ವನ್ನು ದರಪಟ್ಟಿಯಲ್ಲಿ ಕಡ್ಡಾಯವಾಗಿ ನಮೂದಿಸತಕ್ಕದ್ದು.
6. ಮೂಲ ತಯಾರಕರಿಂದ ಪಡೆದ ಪರವಾನಿಗೆ ಪತ್ರ
7. ಸೀಲು ಮಾಡಿದ ಲಕೋಟೆಯ ಮೇಲೆ 'Quotation for the Supply of "Equipments for COVID-19 treatment use in BIMS Hospital, Belagavi" ತಪ್ಪದೆ ನಮೂದಿಸತಕ್ಕದ್ದು.



ನಿರ್ದೇಶಕರು,

ಬೆಳಗಾವಿ ವೈದ್ಯಕೀಯ ವಿಜ್ಞಾನ ಸಂಸ್ಥೆ,
ಬೆಳಗಾವಿ

ಪ್ರತಿಯನ್ನು ಈ ಕೆಳಕಂಡ ಅಧಿಕಾರಿಗಳ ಮಾಹಿತಿಗಾಗಿ ಹಾಗೂ ವ್ಯಾಪಕ ಪ್ರಚಾರಕ್ಕಾಗಿ ಸಲ್ಲಿಸಲಾಗಿದೆ

1. ನಿರ್ದೇಶಕರು, ವೈದ್ಯಕೀಯ ಶಿಕ್ಷಣ, ಆನಂದರಾವ್ ವೃತ್ತ, ಬೆಂಗಳೂರು.
2. ನಿರ್ದೇಶಕರು, ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಸೇವೆಗಳು, ಆನಂದರಾವ್ ವೃತ್ತ, ಬೆಂಗಳೂರು.
3. ಜಿಲ್ಲಾಧಿಕಾರಿಗಳು, ಬೆಳಗಾವಿ ಜಿಲ್ಲೆ, ಬೆಳಗಾವಿ.
4. ಮುಖ್ಯ ಕಾರ್ಯನಿರ್ವಾಹಕ ಅಧಿಕಾರಿಗಳು, ಜಿಲ್ಲಾ ಪಂಚಾಯತ್ ಬೆಳಗಾವಿ.
5. ಜಿಲ್ಲಾ ಆರೋಗ್ಯ ಮತ್ತು ಕುಟುಂಬ ಕಲ್ಯಾಣ ಅಧಿಕಾರಿಗಳು, ಬೆಳಗಾವಿ.
6. ಕಾರ್ಯಪಾಲಕ ಅಭಿಯಂತರರು, ಲೋಕೋಪಯೋಗಿ ಇಲಾಖೆ, ಬೆಳಗಾವಿ.
7. ನಿರ್ದೇಶಕರು /ಬಿ.ಎಮ್.ಸಿ.ಆರ್.ಐ. ಬೆಂಗಳೂರು /ಎಮ್.ಎಮ್.ಸಿ.ಆರ್.ಐ. ಮೈಸೂರು/ ಬಳ್ಳಾರಿ/ಹುಬ್ಬಳ್ಳಿ/ ಡೀನ್ ಶಿವಮೊಗ್ಗ/ ಮಂಡ್ಯ/ ಬೀದರ/ ಹಾಸನ/ ರಾಯಚೂರು/ ಕೊಪ್ಪಳ/ ಗದಗ/ ಕೊಡಗು/ ಗುಲ್ಬರ್ಗಾ/ ಕಾರವಾರ/ ಚಾಮರಾಜನಗರ/ ಜಯದೇವ ಹೃದ್ರೋಗ ಆಸ್ಪತ್ರೆ ಬೆಂಗಳೂರು
8. ಜಿಲ್ಲಾ ಶಸ್ತ್ರ ಚಿಕಿತ್ಸಕರು ಹಾಗೂ ವೈದ್ಯಕೀಯ ಅಧೀಕ್ಷಕರು, ಧಾರವಾಡ/ ವಿಜಯಪುರ/ ಹಾವೇರಿ/ ಬಾಗಲಕೋಟೆ/ ಚಿತ್ರದುರ್ಗ/ ದಾವಣಗೆರೆ.
9. ಸೂಚನಾ ಫಲಕ ಬಿಮ್ಸ್/ಬಿಮ್ಸ್ ಆಸ್ಪತ್ರೆ, ಬೆಳಗಾವಿ.